Bios GmbH Labordiagnostik, München
www.bios-world.com

Test evaluation with patient sera

Sensitivity Data

Biognost® IFA Assays

Parameter:	Mumps IgM

Parameter Description: IgM antibodies to Mumps Virus

lower reference value:	**0**
number of sera tested:	**4**
standard deviation:	**886**
medium titer:	**720**
highest titer:	**2048**
number of sera < reference range:	**0**
sensitivity:	**100%**
data per:	**21.09.2020**

Patient sera are received:

a) from renowned American blood banks, which supply us patients sera for positive controls;

b) from institutions inside and outside Germany, which send in sera for double checking of their results;

c) from Reference Laboratories (WHO, NIBSC, PEI etc.) inside and outside Germany, from which we purchase reference sera for test evaluation or which otherwise make charakterized sera availible to our enterprise.

All data listed have been obtained with single sera.