

Test evaluation with blood donor sera

Specificity Data

Biognost® IFA Assays

Parameter:	ANAHEp
------------	--------

Parameter description:	Nuclear antibodies on Hep-2 cells
------------------------	-----------------------------------

upper reference value:	67
number of sera tested:	569
standard deviation:	28
medium titer*:	12
highest titer:	256
lowest titer:	1
number of sera > reference range:	8
specificity	99%
Data per:	11.09.2023

Sera tested are taken from students and researchers from München universities as well as teachers from the Frauen-Computer-Schule AG München, coming from all parts of Germany as well as neighbour countries (Austria, Switzerland, France, Italy, Holland etc.).

Blood is taken equally from female and male blood donors. The age of our donors persons ranks from 19 - 55 with a strong focus on 20 - 30 years.

*Threshold titer, "Durchseuchungstiter"
